

### CONNECTED VARIABLE VOLUME ELECTRONIC PIPETTES



#### PIPETMAN M SINGLE CHANNEL — VARIABLE VOLUME MODELS

Model	PIPETMAN DIAMOND Tips		Part Number	Gilson Specifications				ISO 8655-2		
				Standard PIPET mode				REPETITIVE mode volume	Systematic error (µL)	Random error (µL)
				Volume Range	Volume (µL)	Systematic error (µL)	Random error (µL)			
P10M	D10* DL10*	DF10ST DFL10ST	<b>F81022</b>	0.5-10 µL	0.5 1 5 10	± 0.040 ± 0.025 ± 0.060 ± 0.080	≤ 0.013 ≤ 0.012 ≤ 0.020 ≤ 0.025	0.5-10 µL	± 0.12 ± 0.12 ± 0.12 ± 0.12	≤ 0.08 ≤ 0.08 ≤ 0.08 ≤ 0.08
P20M	D200	DF30ST	<b>F81023</b>	2-20 µL	2 10 20	± 0.075 ± 0.100 ± 0.150	≤ 0.025 ≤ 0.035 ≤ 0.050	2-20 µL	± 0.2 ± 0.2 ± 0.2	≤ 0.1 ≤ 0.1 ≤ 0.1
P100M	D200	DF100ST	<b>F81013</b>	5-100 µL	5 10 50 100	± 0.35 ± 0.30 ± 0.38 ± 0.40	≤ 0.10 ≤ 0.10 ≤ 0.12 ≤ 0.15	5-100 µL	± 0.8 ± 0.8 ± 0.8 ± 0.8	≤ 0.3 ≤ 0.3 ≤ 0.3 ≤ 0.3
P200M	D200 D300	DF200ST DF300ST	<b>F81024</b>	20-200 µL	20 100 200	± 0.40 ± 0.80 ± 1.00	≤ 0.15 ≤ 0.22 ≤ 0.26	5 - 200 µL	± 1.6 ± 1.6 ± 1.6	≤ 0.6 ≤ 0.6 ≤ 0.6
P300M	D200** D300	DF200ST** DF300ST	<b>F81014</b>	20-300 µL	20 30 150 300	± 0.80 ± 0.70 ± 0.90 ± 1.05	≤ 0.16 ≤ 0.20 ≤ 0.23 ≤ 0.30	10-300 µL	± 4.0 ± 4.0 ± 4.0 ± 4.0	≤ 1.5 ≤ 1.5 ≤ 1.5 ≤ 1.5
P1200M	D1000** D1200	DF1000ST** DF1200ST	<b>F81015</b>	100-1200 µL	100 120 600 1200	± 2.5 ± 2.4 ± 3.6 ± 6.0	≤ 0.4 ≤ 0.4 ≤ 0.8 ≤ 1.2	20-1200 µL	± 16 ± 16 ± 16 ± 16	≤ 6.0 ≤ 6.0 ≤ 6.0 ≤ 6.0
P5000M	D5000		<b>F81016</b>	500-5000 µL	500 2500 5000	± 10 ± 15 ± 25	≤ 2 ≤ 4 ≤ 7	100-5000 µL	± 40 ± 40 ± 40	≤ 15.0 ≤ 15.0 ≤ 15.0
P10mLM	D10mL		<b>F81017</b>	1-10 mL	1 mL 5 mL 10 mL	± 25 ± 30 ± 50	≤ 4 ≤ 8 ≤ 12	200 µL-10 mL	± 60 ± 60 ± 60	≤ 30.0 ≤ 30.0 ≤ 30.0

#### PIPETMAN M MULTICHANNEL

P8x10M	D10 DL10	DF10ST DFL10ST	<b>F81025</b>	0.5-10 µL	0.5 1 5 10	± 0.05 ± 0.04 ± 0.08 ± 0.10	≤ 0.02 ≤ 0.02 ≤ 0.04 ≤ 0.06	0.5-10 µL	± 0.24 ± 0.24 ± 0.24 ± 0.24	≤ 0.16 ≤ 0.16 ≤ 0.16 ≤ 0.16
P12x10M			<b>F81026</b>							
P8x20M	DL10 D200	DFL10ST DF30ST	<b>F81027</b>	1-20 µL	1 2 10 20	± 0.08 ± 0.09 ± 0.15 ± 0.25	≤ 0.05 ≤ 0.06 ≤ 0.10 ≤ 0.12	1-20 µL	± 0.4 ± 0.4 ± 0.4 ± 0.4	≤ 0.2 ≤ 0.2 ≤ 0.2 ≤ 0.2
P12x20M			<b>F81028</b>							
P8x100M	D200	DF100ST	<b>F81018</b>	10-100 µL	10 50 100	± 0.25 ± 0.50 ± 0.80	≤ 0.14 ≤ 0.20 ≤ 0.25	5-100 µL	± 1.6 ± 1.6 ± 1.6	≤ 0.6 ≤ 0.6 ≤ 0.6
P12x100M			<b>F81019</b>							
P8x200M	D200 D300	DF100ST** DF200ST DF300ST	<b>F81029</b>	20-200 µL	20 100 200	± 0.50 ± 1.00 ± 2.00	≤ 0.16 ≤ 0.30 ≤ 0.50	5-200 µL	± 3.2 ± 3.2 ± 3.2	≤ 1.2 ≤ 1.2 ≤ 1.2
P12x200M			<b>F81030</b>							
P8x300M	D200** D300	DF200ST** DF300ST	<b>F81031</b>	10-300 µL	10 30 150 300	± 1.00 ± 1.00 ± 1.50 ± 2.40	≤ 0.18 ≤ 0.18 ≤ 0.375 ≤ 0.45	10-300 µL	± 8.0 ± 8.0 ± 8.0 ± 8.0	≤ 3.0 ≤ 3.0 ≤ 3.0 ≤ 3.0
P12x300M			<b>F81032</b>							
P8x1200M	D1200	DF1200ST	<b>F81020</b>	50-1200 µL	50 120 600 1200	± 4.0 ± 4.0 ± 6.0 ± 9.6	≤ 0.7 ≤ 0.7 ≤ 1.5 ≤ 1.8	50-1200 µL	± 32 ± 32 ± 32 ± 32	≤ 12 ≤ 12 ≤ 12 ≤ 12
P12x1200M			<b>F81021</b>							

Gilson maximum permissible errors are guaranteed only when PIPETMAN pipettes are used with the recommended PIPETMAN DIAMOND Tips.

Under these conditions, Gilson volumetric specifications in standard pipetting (PIPET mode) are guaranteed with a performance exceeding ISO 8655-2 recommendations for this mode.

\* A plastic adapter is required to eject D10 and DF10ST tips (short tips). The adapter is supplied with P10M. No adapter is required when using DL10 and DFL10ST tips (long tips).

\*\* These PIPETMAN DIAMOND Tips can be used with the indicated pipettes for the REPETITIVE mode but not until the maximal volume of the pipette in PIPET mode. Please refer to the volume range of your tips.

Recommended tips to get best results in REPETITIVE mode.

